## **Tanning Device Timer & Emergency Button Test**

This test is required to be completed minimally once a year for each UV Tanning Device.

Timers must be accurate within 10%. If the timer does not pass, take the tanning bed out of service until the timer has been repaired.

NOTE: 10% of 2 minutes is 12 seconds.

Test the emergency shut off button located on the tanning device. If the emergency button fails (does not work), take the tanning bed out of service until it has been repaired.

## KEEP THIS RECORD WITHIN THE FACILITY FOR INSPECTOR REVIEW

Date	<b>Employee Name</b>	Tanning Room Number	Time Set For	Recorded Time	Timer Pass/Fail	Emergency Shut Off Pass/Fail
<u>Example</u> MM/DD/YYYY	First & Last Name	Rm 1	2:00 min.	2:03	Pass	Pass

Date	Employee Name	Tanning Room Number	Time Set For	Recorded Time	Timer Pass/Fail	Emergency Shut Off Pass/Fail